Establishment Name:	Kate's Kitchen	
<b>Establishment Address:</b>	1234 Blonde Ave.	

City / State / Zip: Des Moines, IA 50319

Phone: (111) 111-1111

## Criteria:

1. Closely follow manufacturer's instructions for proper calibration of pH meter(s).

2. Record pH in two decimal points.

_	Calibration Buffer	Buffer Code	<b>Expiration Date</b>	Buffer Notes				
pH Buffer #1:	4.01	xyz123	10/01/2023	Into service on 08/25/2022				
pH Buffer #2:	7.00	tqy555	09/01/2022	Into service 07/15/2021. Re-ordered 07/01/2022				
pH Buffer #3:	10.00	Not Used						
pH Meter ID: Unit #1: Serial Number CP5772CV325								
pH Meter ID:	pH Meter ID: Unit #2: Serial Number CP5883XTZ525							

			Buffer #1	Buffer #2	Buffer #3		Operator	Operating	
Date	Meter ID	Time	Results	Results	Results	Temp.	Initials	Properly? Y/N	Calibration Notes
08/21/2022	#1	0830	4.00	7.00		77.6 F	JK	Yes	4.00 = within accuracy spec of tester
08/21/2022	#2	0915	4.00	Error		76.3 F	JK	No	Tester malfunctioned - taken out of service
08/22/2022	#1	1015	4.01	7.00		78.1 F	JK	Yes	
09/01/2022	#2	0930	4.00	7.00		77.1 F	JK	Yes	Tester repaired and functioning properly. Out of pH 7.00 Buffer
									End of this report. See next Calibration Log with new buffer supply